



National Accreditation Board for
Testing and Calibration Laboratories

CERTIFICATE OF ACCREDITATION

**YOUNG ENGG & CALIBRATION SERVICES PRIVATE
LIMITED**

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2017

**"General Requirements for the Competence of Testing &
Calibration Laboratories"**

for its facilities at

B20, SAUJANYA SOCIETY, VADODARA, GUJARAT, INDIA

in the field of

CALIBRATION

Certificate Number: CC-4241

Issue Date: 26/01/2025

Valid Until: 25/01/2029

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.

(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Name of Legal Entity: YOUNG ENGG & CALIBRATION SERVICES PRIVATE LIMITED

Signed for and on behalf of NABL



Anita Rani
Director

N. Venkateswaran
Chief Executive Officer



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

YOUNG ENGG & CALIBRATION SERVICES PRIVATE LIMITED, B20, SAUJANYA SOCIETY, VADODARA, GUJARAT, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4241

Page No

1 of 8

Validity

26/01/2025 to 25/01/2029

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz)	AC Current @ 45 Hz to 1 kHz	Using MFC and 50 Turn Coil by Direct method	20 A to 1000 A	4.5 % to 0.52 %
2	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz)	AC Current @ 45 Hz to 1 kHz	Using 6½ DMM by Direct method	30 µA to 10 A	0.7 % to 0.3 %
3	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz)	AC Current @ 50 Hz to 1 kHz	Using MFC by Direct method	30 µA to 10 A	3.2 % to 0.72 %
4	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz)	AC Voltage @ 45 Hz to 1 kHz	Using 6½ DMM by Direct method	1 mV to 1000 V	4.78 % to 0.3 %
5	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz)	AC Voltage @ 45 Hz to 1 kHz	Using MFC by Direct method	1 mV to 1000 V	9.9 % to 0.2 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

YOUNG ENGG & CALIBRATION SERVICES PRIVATE LIMITED, B20, SAUJANYA SOCIETY, VADODARA, GUJARAT, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4241

Page No

2 of 8

Validity

26/01/2025 to 25/01/2029

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
6	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz)	Transformer Turns Ratio	Using 6½ DMM (Two DMM used) & Turns Ratio Calibrator by V/V Method	11 to 1100	0.3 % to 1.4 %
7	ELECTRO-TECHNICAL-DIRECT CURRENT	DC Current	Using 6½ DMM by Direct method	1 mA to 10 A	0.36 % to 0.22 %
8	ELECTRO-TECHNICAL-DIRECT CURRENT	DC Current	Using Multifunction Calibrator by Direct method	1 mA to 10 A	1.48 % to 0.63 %
9	ELECTRO-TECHNICAL-DIRECT CURRENT	DC Current	Using MFC and 50 Turn Coil by Direct method	20 A to 1000 A	4.503 % to 0.49 %
10	ELECTRO-TECHNICAL-DIRECT CURRENT	DC Voltage	Using MFC by Direct method	1 mV to 1000 V	6.22 % to 0.064 %
11	ELECTRO-TECHNICAL-DIRECT CURRENT	DC Voltage	Using 6½ DMM by Direct method	1 mV to 1000 V	0.63 % to 0.01 %
12	ELECTRO-TECHNICAL-DIRECT CURRENT	Resistance (2 Wire)	Using 6½ DMM by Direct method	1 Mohm to 19 Mohm	0.02 % to 2.97 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

YOUNG ENGG & CALIBRATION SERVICES PRIVATE LIMITED, B20, SAUJANYA SOCIETY, VADODARA, GUJARAT, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4241

Page No

3 of 8

Validity

26/01/2025 to 25/01/2029

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
13	ELECTRO-TECHNICAL-DIRECT CURRENT	Resistance (2 Wire)	Using MFC by Direct method	1 ohm to 100 kohm	6.56 % to 0.084 %
14	ELECTRO-TECHNICAL-DIRECT CURRENT	Resistance (2 Wire)	Using Multifunction Calibrator by Direct method	100 kohm to 190 Mohm	0.084 % to 1.21 %
15	ELECTRO-TECHNICAL-DIRECT CURRENT	Resistance (4 Wire)	Using 6½ DMM by Direct method	1 ohm to 1 Mohm	0.7 % to 0.02 %
16	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION	J type Thermocouple	Using 6½ DMM by Direct method	25 °C to 1000 °C	0.6 °C
17	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION	J type Thermocouple	Using Multifunction Calibrator by Direct method	25 °C to 1300 °C	1.32°C
18	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION	K type Thermocouple	Using 6½ DMM by Direct method	200 °C to 1300 °C	0.6 °C
19	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION	K type Thermocouple	Using Multifunction Calibrator by Direct method	25 °C to 1300 °C	1.32 °C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

YOUNG ENGG & CALIBRATION SERVICES PRIVATE LIMITED, B20, SAUJANYA SOCIETY, VADODARA, GUJARAT, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4241

Page No

4 of 8

Validity

26/01/2025 to 25/01/2029

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
20	ELECTRO-TECHNICAL-TIME & FREQUENCY	Frequency	Using DMM by Direct method	45 Hz to 1 kHz	0.22 % to 0.10 %
21	ELECTRO-TECHNICAL-TIME & FREQUENCY	Frequency @ 1 V	Using Multifunction Calibrator by Direct method	45 Hz to 1000 Hz	0.13 % to 0.015 %
22	ELECTRO-TECHNICAL-TIME & FREQUENCY	Time	Using Standard Time Calibrator by Comparison method	1 hr to 24 hr	3.5 s to 50.16 s



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

YOUNG ENGG & CALIBRATION SERVICES PRIVATE LIMITED, B20, SAUJANYA SOCIETY, VADODARA, GUJARAT, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4241

Page No

5 of 8

Validity

26/01/2025 to 25/01/2029

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Site Facility					
1	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz)	AC Current @ 45 Hz to 1 kHz	Using MFC and 50 Turn Coil by Direct method	20 A to 1000 A	4.5 % to 0.52 %
2	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz)	AC Current @ 50 Hz to 1 kHz	Using MFC by Direct method	30 µA to 10 A	3.2 % to 0.72 %
3	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz)	AC High Voltage @ 50 Hz	Using HV Divider with DU by direct method	1 kV to 100 kV	3.5 %
4	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz)	AC Voltage @ 45 Hz to 1 kHz	Using 6½ DMM by Direct method	1 mV to 1000 V	4.78 % to 0.3 %
5	ELECTRO-TECHNICAL-Alternating Current (< 1 GHz)	AC Voltage @ 45 Hz to 1 kHz	Using MFC by Direct method	1 mV to 1000 V	9.9 % to 0.2 %
6	ELECTRO-TECHNICAL-DIRECT CURRENT	DC Current	Using 6½ DMM by Direct method	1 mA to 10 A	0.36 % to 0.22 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

YOUNG ENGG & CALIBRATION SERVICES PRIVATE LIMITED, B20, SAUJANYA SOCIETY, VADODARA, GUJARAT, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4241

Page No

6 of 8

Validity

26/01/2025 to 25/01/2029

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured / Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
7	ELECTRO-TECHNICAL-DIRECT CURRENT	DC Current	Using Multifunction Calibrator by Direct method	1 mA to 10 A	1.48 % to 0.63 %
8	ELECTRO-TECHNICAL-DIRECT CURRENT	DC Current	Using MFC and 50 Turn Coil by Direct method	20 A to 1000 A	4.503 % to 0.49 %
9	ELECTRO-TECHNICAL-DIRECT CURRENT	DC Voltage	Using MFC by Direct method	1 mV to 1000 V	6.22 % to 0.064 %
10	ELECTRO-TECHNICAL-DIRECT CURRENT	DC Voltage	Using 6½ DMM by Direct method	1 mV to 1000 V	0.63 % to 0.01 %
11	ELECTRO-TECHNICAL-DIRECT CURRENT	Resistance (2 Wire)	Using 6½ DMM by Direct method	1 Mohm to 19 Mohm	0.02 % to 2.97 %
12	ELECTRO-TECHNICAL-DIRECT CURRENT	Resistance (2 Wire)	Using MFC by Direct method	1 ohm to 100 kohm	6.56 % to 0.084 %
13	ELECTRO-TECHNICAL-DIRECT CURRENT	Resistance (2 Wire)	Using Multifunction Calibrator by Direct method	100 kohm to 190 Mohm	0.084 % to 1.21 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

YOUNG ENGG & CALIBRATION SERVICES PRIVATE LIMITED, B20, SAUJANYA SOCIETY, VADODARA, GUJARAT, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4241

Page No

7 of 8

Validity

26/01/2025 to 25/01/2029

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
14	ELECTRO-TECHNICAL-DIRECT CURRENT	Resistance (4 Wire)	Using 6½ DMM by Direct method	1 ohm to 1 Mohm	0.7 % to 0.02 %
15	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION	J type Thermocouple	Using 6½ DMM by Direct method	25 °C to 1000 °C	0.6 °C
16	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION	J type Thermocouple	Using Multifunction Calibrator by Direct method	25 °C to 1300 °C	1.32°C
17	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION	K type Thermocouple	Using 6½ DMM by Direct method	200 °C to 1300 °C	0.6 °C
18	ELECTRO-TECHNICAL-TEMPERATURE SIMULATION	K type Thermocouple	Using Multifunction Calibrator by Direct method	25 °C to 1300 °C	1.32 °C
19	ELECTRO-TECHNICAL-TIME & FREQUENCY	Frequency	Using DMM by Direct method	45 Hz to 1 kHz	0.22 % to 0.10 %
20	ELECTRO-TECHNICAL-TIME & FREQUENCY	Frequency @ 1 V	Using Multifunction Calibrator by Direct method	45 Hz to 1000 Hz	0.13 % to 0.015 %



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

YOUNG ENGG & CALIBRATION SERVICES PRIVATE LIMITED, B20, SAUJANYA SOCIETY, VADODARA, GUJARAT, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-4241

Page No

8 of 8

Validity

26/01/2025 to 25/01/2029

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
21	ELECTRO-TECHNICAL-TIME & FREQUENCY	Time	Using Standard Time Calibrator by Comparison method	1 hr to 24 hr	3.5 s to 50.16 s

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.